



## DECLARATION OF CONFORMITY

Manufacturer:

**LIW Care Technology Sp. z o.o.**  
**ul. Golfowa 7**  
**94-406 Łódź, Poland**

Hereby declares that

### **PEDIATRIC PUSHCHAIR BAFFIN BUGGY**

bearing CE mark is a Class I medical device, Rule 1 in accordance with Annex 9 of European Medical Devices Directive 93/42/EEC + 2007/47/EC and fulfills the basic requirements specified in Annex 1 of the Directive.

The conformity assessment was done according to Annex 7 of Directive 93/42/EEC.

Medical device is in conformity with the following harmonized standards:

- **PN-EN 12182:2012**
- **PN-EN ISO 14971:2012**
- **PN-EN 1041+A1:2013-12**
- **PN-EN ISO 15223-1:2012**

On behalf of the manufacturer:  
Tomasz Chmielecki, CEO

Signature:

  
**PREZES ZARZĄDU**

Manufacturer's seal:

Łódź, 16 December 2015

**LIW CARE TECHNOLOGY Sp. z o.o.**  
**94-406 Łódź, ul. Golfowa 7**  
**NIP: 729-266-53-87, REG. 100715121**  
**KRS:0000333719**